



Certificate of Analysis

Customer Information

Client:
Attention:
Address:

Testing Facility

Lab: Cora Science, LLC
Address 8000 Anderson Square, STE 113
Austin, Texas 78757
Contact: info@corascience.com
(512) 856-5007

Sample Image(s)



Sample Information

Name: MIT THERAPY CINNA MAENG DA
Lot Number: MT1002-C
Description: Liquid botanical extract
Condition: Good
Job ID: ISO06695
Sample ID: I18703
Received: 30MAR2026
Completed: 06APR2026
Issued: 07APR2026

Test Results

Mitragyna Alkaloids (UHPLC-DAD)		Method Code: T102		Tested: 02APR2026 0207		
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES	
Mitragynine	Report Results	0.512	w/w%	0.00030	N/A	
7-Hydroxymitragynine	Report Results	0.000460	w/w%	0.00030	N/A	
Paynantheine	Report Results	0.0860	w/w%	0.00030	N/A	
Speciogynine	Report Results	0.0613	w/w%	0.00030	N/A	
Speciociliatine	Report Results	0.118	w/w%	0.00030	N/A	
Total Mitragyna Alkaloids	Report Results	0.778	w/w%	0.00030	N/A	

Mitragyna Alkaloids (UHPLC-DAD)		Method Code: T102		Tested: 02APR2026 0207		
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES	
Mitragynine	Report Results	80.2	mg/unit	0.047	N/A	
7-Hydroxymitragynine	Report Results	0.0721	mg/unit	0.047	N/A	
Paynantheine	Report Results	13.5	mg/unit	0.047	N/A	
Speciogynine	Report Results	9.60	mg/unit	0.047	N/A	
Speciociliatine	Report Results	18.5	mg/unit	0.047	N/A	
Total Mitragyna Alkaloids	Report Results	122	mg/unit	0.047	N/A	

Elemental Impurities (ICP-MS)		Method Code: T301		Tested: 01APR2026 0857		
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES	
Arsenic	NMT 1.50	<LOQ	ug/g	0.006	PASS	
Cadmium	NMT 0.50	<LOQ	ug/g	0.002	PASS	
Mercury	NMT 0.20	<LOQ	ug/g	0.002	PASS	
Lead	NMT 0.50	0.006	ug/g	0.002	PASS	

Residual Solvents: Class I (GC-MS)		Method Code: T201		Tested: 06APR2026 1638	
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES
1,1-Dichloroethene	NMT 8	<LOQ	ug/g	0.40	PASS
1,1,1-Trichloroethane	NMT 1500	<LOQ	ug/g	75	PASS
Tetrachloromethane	NMT 4	<LOQ	ug/g	0.20	PASS
Benzene	NMT 2	<LOQ	ug/g	0.10	PASS
1,2-Dichloroethane	NMT 5	<LOQ	ug/g	0.25	PASS

Residual Solvents: Class II (GC-MS)		Method Code: T201		Tested: 06APR2026 1638	
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES
Methanol	NMT 3000	<LOQ	ug/g	75	PASS
Acetonitrile	NMT 410	<LOQ	ug/g	41	PASS
Dichloromethane	NMT 600	<LOQ	ug/g	15	PASS
1,2-Dichloroethene, (E)	NMT 1870	<LOQ	ug/g	46.75	PASS
1,2-Dichloroethene, (Z)	NMT 1870	<LOQ	ug/g	46.75	PASS
Tetrahydrofuran	NMT 720	<LOQ	ug/g	18	PASS
Cyclohexane	NMT 3880	<LOQ	ug/g	97	PASS
Methylcyclohexane	NMT 1180	<LOQ	ug/g	29.5	PASS
1,4-Dioxane	NMT 380	<LOQ	ug/g	38	PASS
Toluene	NMT 890	<LOQ	ug/g	22.25	PASS
Chlorobenzene	NMT 360	<LOQ	ug/g	9	PASS
Ethylbenzene	NMT 2170	<LOQ	ug/g	54.25	PASS
o/p-Xylene	NMT 2170	<LOQ	ug/g	54.25	PASS
m-Xylene	NMT 2170	<LOQ	ug/g	54.25	PASS
Isopropylbenzene	NMT 70	<LOQ	ug/g	1.75	PASS
Hexane	NMT 290	<LOQ	ug/g	7.25	PASS
Nitromethane	NMT 50	<LOQ	ug/g	1.25	PASS
Chloroform	NMT 60	<LOQ	ug/g	1.5	PASS
1,2-Dimethoxyethane	NMT 100	<LOQ	ug/g	2.5	PASS
Trichloroethene	NMT 80	<LOQ	ug/g	2	PASS
Pyridine	NMT 200	<LOQ	ug/g	5	PASS
2-Hexanone	NMT 50	<LOQ	ug/g	5	PASS
Tetralin	NMT 100	<LOQ	ug/g	2.5	PASS
Total Xylenes	NMT 2170	<LOQ	ug/g	54	PASS

Residual Solvents: Class III (GC-MS)		Method Code: T201		Tested: 06APR2026 1638	
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES
Pentane	NMT 5000	<LOQ	ug/g	125	PASS
Ethanol	NMT 5000	<LOQ	ug/g	125	PASS
Diethyl Ether	NMT 5000	<LOQ	ug/g	125	PASS
Acetone	NMT 5000	<LOQ	ug/g	125	PASS
Ethyl Formate	NMT 5000	<LOQ	ug/g	125	PASS
Isopropanol	NMT 5000	<LOQ	ug/g	125	PASS
Methyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS
Methyl tert-Butyl Ether	NMT 5000	<LOQ	ug/g	125	PASS
1-Propanol	NMT 5000	<LOQ	ug/g	125	PASS

PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES
2-Butanone	NMT 5000	<LOQ	ug/g	125	PASS
Ethyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS
2-Butanol	NMT 5000	<LOQ	ug/g	125	PASS
2-Methyl-1-Propanol	NMT 5000	<LOQ	ug/g	125	PASS
Isopropyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS
Heptane	NMT 5000	<LOQ	ug/g	125	PASS
1-Butanol	NMT 5000	<LOQ	ug/g	125	PASS
Propyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS
4-Methyl-2-Pentanone	NMT 5000	<LOQ	ug/g	125	PASS
Isoamyl Alcohol	NMT 5000	<LOQ	ug/g	125	PASS
Isobutyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS
1-Pentanol	NMT 5000	<LOQ	ug/g	125	PASS
Butyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS
Anisole	NMT 5000	<LOQ	ug/g	125	PASS
Dimethylsulfoxide	NMT 5000	<LOQ	ug/g	125	PASS

Microbial Examination		Method Code: T005		Tested: 02APR2026 1439	
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES
Total Aerobic Plate Count	NMT 10,000 CFU/g	<LOQ	CFU/g	10 CFU/g	PASS
Total Yeast and Mold	NMT 1,000 CFU/g	<LOQ	CFU/g	10 CFU/g	PASS
Total Coliforms	NMT 100 CFU/g	<LOQ	CFU/g	10 CFU/g	PASS
Escherichia coli	Not Detected in 10 g	Not Detected	N/A	1 CFU/10g	PASS
Salmonella spp.	Not Detected in 10 g	Not Detected	N/A	1 CFU/10g	PASS

Loss on Drying		Method Code: T505		Tested: 31MAR2026 1516	
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES
Moisture	Report Results	85.6	%	0.1	N/A

7-Hydroxymitragynine Limit (0.04%)		Method Code: 813		Tested: 02APR2026 0207	
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES
7-Hydroxymitragynine	NMT 400 PPM	32	ppm	21	PASS

Additional Report Notes

T102 result, LOQ and unit converted from w/w% to mg/mL using a laboratory measured density of 1.044 g/mL and package specified fill volume of 15.0 mL. T813 results are reported on a dry-weight basis (DWB). Reported values converted from T102 results using the laboratory-measured loss on drying by T505 for each sample:
DWB w/w% = (as-received w/w%) ÷ (1 – moisture%/100).

Revision History


Report ID: cfa03a2f-92be-4932-b62e-9a32b554ee08
rev 00 - Initial release.
rev 01 - Added T201 results.

Abbreviations

ID: identification, **N/A:** not applicable, **LOQ:** limit of quantitation, **CFU:** colony forming units, **w/w%:** weight by weight percent, **mg:** milligrams, **g:** grams, **ug:** micrograms, **mL:** milliliters, **ND:** not detected, **<LOQ:** below limit of quantitation, **NMT:** no more than, **NLT:** no less than, **UHPLC:** ultra-high performance liquid chromatography, **GC:** gas chromatography, **DAD:** diode array detection/detector, **MS:** mass spectroscopy/spectrometer, **ICP:** inductively coupled plasma, **ISO:** International Organization for Standardization, **USP:** United States Pharmacopeia

Authorization

This report has been authorized for release from Cora Science by:

Signature:		Position:	Laboratory Director
Name:	Tyler West	Department:	Management
		Date:	07APR2026